



Instruction for Use

SDMA ELISA

Enzyme Immunoassay
for the Quantitative Determination of
Endogenous Symmetric Dimethylarginine (SDMA)
in Serum or Plasma



REF EA203/96



12 x 8



2 – 8 °C

DLD Gesellschaft für Diagnostika und medizinische Geräte mbH
Adlerhorst 15 • 22459 Hamburg • Germany
Tel +49-40-555 87 10 • Fax +49-40-555 87 111
Internet: <http://www.dld-diagnostika.de> • E-Mail: contact@dld-diagnostika.de

Contents

1. Introduction and Principle of the Test	Page 4
2. Precautions	Page 5
3. Storage and Stability	Page 5
4. Contents of the Kit	Page 5
5. Sample Collection	Page 7
6. Preparation of Reagents and Samples	Page 7
7. Test Procedure (ELISA)	Page 10
8. Calculation of the Results	Page 11
9. Assay Characteristics	Page 12
10. Literature	Page 14
Pipetting Scheme Sample Preparation	Page 15
Pipetting Scheme ELISA	Page 16

1. Introduction and Principle of the Test

Dosing of most drugs must be adapted in renal insufficiency, making accurate assessment of renal function an essential component of diagnostics in clinical medicine. Furthermore, even modest impairment of renal function has been recognized as a cardiovascular risk factor. As the most commonly used marker of renal excretory function, serum creatinine concentration, does not adequately respond to mild to moderate impairment of renal function, more sensitive markers for renal excretory function are urgently sought, especially in mild stages of renal impairment. SDMA is a methylated derivative of the amino acid L-arginine (symmetric dimethylarginine). SDMA is eliminated from the body exclusively by renal excretion; therefore SDMA plasma concentration is tightly related to renal function. Thus, quantification of plasma SDMA is an adequate means to assess renal function, as could be demonstrated in a series of recent clinical trials: In 18 clinical studies involving more than 2,100 patients systemic SDMA concentrations were highly correlated with inulin clearance as well as with various clearance estimates and better corresponded to mild renal function impairment than serum creatinine.

Thus, SDMA exhibits properties of a reliable marker of renal function. Furthermore, there is evidence showing that elevated SDMA levels, as they may occur in renal function impairment, may prospectively indicate future risk of cardiovascular disease and mortality independently of the level of renal impairment.

The competitive SDMA-ELISA uses the microtiter plate format. SDMA is bound to the solid phase of the microtiter plate. SDMA in the samples is acylated and competes with solid phase bound SDMA for a fixed number of rabbit anti-SDMA antiserum binding sites. When the system is in equilibrium, free antigen and free antigen-antiserum complexes are removed by washing. The antibody bound to the solid phase SDMA is detected by anti-rabbit / peroxidase. The substrate TMB / peroxidase reaction is monitored at 450 nm. The amount of antibody bound to the solid phase SDMA is inversely proportional to the SDMA concentration of the sample.

2. Precautions

- For in vitro use only.
- Disposable gloves should be used.
- Material of animal origin used in the preparation of the kit has been obtained from animals certified as healthy but these materials should be handled as potentially infectious.

3. Storage and Stability

On arrival, store the kit at 2-8 °C. Once opened the kit is stable until its expiry date. For stability of prepared reagents refer to Preparation of Reagents.

Do not use components beyond the expiration date shown on the labels.
Do not mix various lots of any kit component within an individual assay.

4. Contents of the Kit

4.1 **MT-Strips** **STRIPS** 12 strips
8 wells each, break apart
precoated with SDMA

4.2 **Standards 1 - 6** **CAL 1 – 6** 6 vials
Each 4 ml, ready for use
Concentrations:

Standard	1	2	3	4	5	6
µmol/l	0	0.2	0.4	0.7	1.2	3.0
ng/ml	0	40	81	141	242	606

4.3 **Control 1 & 2** **CON 1 & 2** 2 vials
Each 4 ml, ready for use
Range: see q.c. certificate

4.4 **Acylation Buffer** **ACYL-BUFF** 1 vial
3.5 ml, ready for use, blue coloured

4.5 **Acylation Reagent** **ACYL-REAG** 3 vials
lyophilised, dissolve contents
in 3 ml Solvent before use

4.6	Antiserum 7 ml, ready for use yellow coloured, Rabbit-anti-N-acyl-SDMA	AS	1 vial
4.7	Enzyme Conjugate 13 ml, ready for use goat anti-rabbit-IgG-peroxidase	CONJ	1 vial
4.8	Wash Buffer 20 ml, 50 x concentrated Dilute contents with dist. water to 1000 ml total volume.	WASH	1 vial
4.9	Substrate 13 ml TMB solution, ready for use	SUB	1 vial
4.10	Stop Solution 13 ml, ready for use Contains 0.3 M sulphuric acid, not corrosive	STOP	1 vial
4.11	Reaction Plate for acylation	ACYL-PLATE	1 piece
4.12	Equalizing Reagent lyophilised, dissolve contents with 21 ml dist. water, dissolve carefully to minimize foam formation	EQUA-REAG	1 vial
4.13	Solvent 5 ml, contains DMSO Please note that Solvent reacts with many plastic materials including plastic trays; Solvent does not react with normal pipette tips and with glass devices	SOLVENT	2 vials
4.14	Foil	FOIL	2 pieces

Additional materials and equipment required but not provided:

- Pipettes (20, 50, 100 and 200 µl,)
- Multipette
- Orbital shaker
- Microplate washing device
- Microplate photometer (450 nm)
- Vortex mixer, roll mixer

5. Sample Collection

5.1. Serum and Plasma

The test can be performed with serum as well as with EDTA plasma.

Hemolytic and lipemic samples should not be used.

The samples can be stored up to 6 hours at 2 - 8 °C. For a longer storage (up to 18 months) the samples must be kept frozen at -20 °C

Repeated freezing and thawing should be avoided.

6. Preparation of Reagents and Samples

6.1. Microtiter strips **STRIPS**

Before opening the packet of strip wells, allow it to stand at room temperature for at least 10 minutes. After opening, keep any unused wells in the original foil packet with the desiccant provided. Reseal carefully and store at 2-8 °C.

6.2 Wash Buffer **WASH**

Dilute the contents with dist. water to a total volume of 1000 ml, mix shortly. The diluted wash buffer must be stored at 2 - 8 °C and is stable for 4 weeks. For longer storage the diluted wash buffer has to be stored frozen at -20°C.

6.3. Equalizing Reagent **EQUA-REAG**

Dissolve the contents with 21 ml dist. water, mix shortly and leave on a roll mixer for 20 minutes. Avoid excess formation of foam. The reconstituted Equalizing Reagent should be stored frozen at -20 °C and is stable until expiry date.

6.4. Acylation Reagent **ACYL-REAG**

Dissolve the contents of one bottle in 3 ml Solvent and shake for 10 minutes on an orbital shaker. After use the reagent has to be discarded. The Acylation Reagent has always to be prepared immediately before use and is stable for minimum 3 hours. The two other bottles allow a second and third run of the test. If the whole kit is to be used in one run it

is recommended to pool the dissolved contents of the two vials of Acylation Reagent.

Please note that Solvent reacts with many plastic materials including plastic trays which are used as reservoir for multichannel pipettes. Solvent does not react with normal pipette tips and with glass devices. It is recommended to use a multipipette, fill it directly from the vial and add the Acylation Reagent to the wells.

All other reagents are ready for use.

6.5. Preparation of Samples (Acylation)

The wells of the reaction plate for the acylation can be used only once. Please mark the respective wells before use to avoid repeated use.

1. Pipette each 20 µl standard 1 - 6, each 20 µl control 1 & 2 and each 20 µl patient sample into the respective wells of the Reaction Plate.
2. Pipette 20 µl Acylation Buffer into all wells.
3. Pipette 200 µl reconstituted Equalizing Reagent into all wells.
4. Mix the reaction plate for 10 seconds.
5. Prepare Acylation Reagent freshly and pipette 50 µl prepared Acylation Reagent each into all wells, mix immediately.
It is recommended to use a multipipette, fill it directly from the vial and add the Acylation Reagent to the wells.
Colour changes to violet.
6. Incubate for 20 minutes at room temperature (approx. 20 °C) on an orbital shaker. Do not cover wells or plate, leave the plate open on the shaker.

Take each 20 µl of the acylated samples for the SDMA-ELISA.

7. Test Procedure ELISA

Bring all reagents to room temperature and mix them carefully, avoid development of foam.

7.1 Sample Incubation

Pipette each 20 µl prepared Standards 1 to 6, 20 µl prepared controls and 20 µl prepared samples into the respective wells of the coated microtiter strips (duplicates are recommended).

Pipette each 50 µl Antiserum into all wells.

Cover the plate with adhesive foil and incubate Microtiter Strips for 90 minutes at room temperature (20 - 25°C) on an orbital shaker.

7.2 Washing

Discard or aspirate the contents of the wells and wash thoroughly with each 300 µl Wash Buffer (shake shortly on an orbital shaker). Repeat the washing procedure 4 times. Remove residual liquid by tapping the inverted plate on clean absorbent paper.

7.3 Conjugate Incubation

Pipette each 100 µl enzyme conjugate into all wells.

Incubate for 30 minutes at room temperature on an orbital shaker.

7.4 Washing

Repeat step 7.2.

7.5 Substrate Incubation

Pipette each 100 µl Substrate into all wells and incubate for 25 ± 5 minutes at room temperature on an orbital shaker.

7.6 Stopping

Pipette each 100 µl Stop Solution into all wells.

7.7 Reading


Read the optical density at 450 nm (reference wavelength between 570 and 650 nm) in a microplate photometer.

8. Calculation of the Results

On a semilogarithmic graph paper the concentration of the standards (x-axis, logarithmic) are plotted against their corresponding optical density (y-axis, linear). Cubic spline, 4 parameter or similar iteration procedures are recommended for evaluation of the standard curve.

The concentration of the controls and samples can be read directly from this standard curve by using their average optical density.

Typical standard curve:



Conversion factor: 1 µmol/l = 202 ng/ml

9. Assay Characteristics

Expected Values

0.30 – 0.75 µmol/l (60 – 150 ng/ml)

The reference ranges given above should only be taken as a guideline. It is recommended that each laboratory should establish its own reference values.

Sensitivity

0.03 µmol/l

Recovery

	Range (µmol/l)	Mean (%)	Range (%)
EDTA-Plasma	0.43 – 1.44	97	86 - 104
Serum	0.45 – 1.72	93	88 - 102

Linearity

	Range (µmol/l)	Highest Dil.	Mean (%)	Range (%)
EDTA-Plasma	0,23 – 1,72	1 : 6 with water	97	89 – 105

Specificity (Cross Reactivity)

Substance	Cross Reactivity (%)
SDMA	100
ADMA	0.74
NMMA	0.76
Homoarginine	0.04
Arginine	0.01

Reproducibility

	Range (µmol/l)	Intra Assay CV
EDTA-Plasma	0.52 – 0.82	6.2 – 4.9 %

	Range (µmol/l)	Inter Assay CV
EDTA-Plasma	0.52 – 1.21	2.0 – 8.8 %

Method Comparison

	Method	Correlation
Serum / Plasma	LC/MS	$Y = 0.96 \times \text{LC/MS} + 0.05$ $R = 0.987; N = 32$

10. Literature

Bode-Böger S.M., Scalera F., Kielstein J.T., Martens-Lobenhoffer J., Breithardt G., Fobker M., Reinecke H.

Symmetrical Dimethylarginine: A new combined parameter for renal function and extent of coronary artery disease

J. Am. Soc. Nephrol. (2006) **17**: 1128-1134

Kielstein J.T., Salpeter S.R.; Bode-Böger S.M., Cooke J.P., Fliser D.

Symmetric dimethylarginine (SDMA) as endogenous marker of renal function – a meta-analysis

Nephrol. Dial. Transplant (2006) **21**: 2446 - 2451

Wanby P., Teerlink T., Brudin L., Brattström L., Nilsson I., Palmqvist P., Carlsson M.

Asymmetric dimethylarginine (ADMA) as a risk marker for stroke and TIA in a Swedish population

Atherosclerosis (2006) **185**: 271 - 277

Pipetting Scheme Sample Preparation

		Standard	Control	Sample
Standard 1 - 6	μl	20		
Control 1 & 2	μl		20	
Patient Sample	μl			20
Acylation Buffer	μl	20	20	20
Equalizing Reagent	μl	200	200	200

shake for 10 seconds

freshly prepared Acylation Reagent	μl	50	50	50
---------------------------------------	----	----	----	----

incubate for 20 minutes at room temperature on an orbital shaker

Pipetting Scheme ELISA

		Standard	Control	Sample
Standard 1 - 6	µl	20		
Control 1 & 2	µl		20	
Patient Sample	µl			20
Antiserum	µl	50	50	50

cover frame with foil and incubate on an orbital shaker
for 90 minutes at room temperature

wash 4 x with each 300 µl Wash Buffer

Enzyme Conjugate	µl	100	100	100
------------------	----	-----	-----	-----

incubate for 30 minutes at room temperature on an orbital shaker

wash 4 x with each 300 µl Wash Buffer

Substrate	µl	100	100	100
-----------	----	-----	-----	-----

incubate for 25 ± 5 minutes at room temperature on an orbital shaker

Stop Solution	µl	100	100	100
---------------	----	-----	-----	-----

read absorbance at 450 nm